Anti-SARS-CoV-2 Spike Protein (Trimer) Human IgA **ELISA Kit**

Summary

Catalog No.	KAV00110
Target	Spike
Stability and Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition.
Detection method	Colorimetric
Sample type	Plasma, Serum
Assay type	Quantitative
Recovery	80-120%
Shipping	2-8 °C
Note	For Research Use Only.

Precision

CV<15%

Data Image



