

HPV18 E6/Protein E6 ELISA Kit

Summary

Catalog No.	KVV09001
Alternative Names	E6, Protein E6
Stability and Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition.
Detection method	Colorimetric
Sample type	Plasma, Serum
Assay type	Quantitative (Sandwich)
Sensitivity	8.72 pg/mL
Range	15.63 - 1,000 pg/mL
Recovery	80-120%
Shipping	2-8 °C
Note	For research use only.

Precision

CV

Data Image
