

Tirzepatide (LY3298176) ELISA Kit-HS

Summary

Catalog No. KDK13903

The stability of ELISA kit is determined by the loss rate of activity. The loss

Stability and Storage rate of this kit is less than 10% prior to the expiration date under

appropriate storage condition.

Detection method Colorimetric

Sample type Plasma, Serum

Assay type Quantitative

Sensitivity 27.45 ng/ml

Range 62.5-4,000 ng/ml

Recovery 80-120%

Shipping 2-8 °C

Note For Research Use Only.

Precision

CV

Data Image